

Eligibility for a Collection Center

1. Technical

Phlebotomist qualification – DMLT or B.Sc. MLT

2. Infrastructure

- a. Air-conditioned Room with minimum area of 150 square feet Reception cum waiting area
- b. Infrastructure should be on ground floor.
- c. Table – Wooden, size 3 x 4 fts
- d. Chairs – 3
- e. Phlebotomy area – minimum size 25 sq ft with partition for privacy
- f. Secure Area – Cabinet with lock and key
- g. Toilet area – minimum 25 sq ft with all the basic amenities.
- h. Electricity Back-up – Inverter 1 KVA
- i. Internet Connection
- j. Dedicated Landline or Telephone connection

3. Stationary Area

- a. Desktop/Laptop
- b. Laser jet / Desk jet printer
- c. Acu-MDx Authorized Collection Center board (As per the company specifications)
- d. Display board for announcements/notices/pictures
- e. Wall Clock

4. Equipment

- a. Centrifuge machine – 8 head with timer with speed 3000 rpm to 3500 rpm.
- b. Weighing Machine (upto 120KG)
- c. 200 Litres Refrigerator for sample storage (No eatable allowed)
- d. Needle Cutter/ Destroyer (electric)
- e. Three color dustbins and bags required for Sharp, for Non-Sharp and General Waste
- f. First aid box
- g. Height Chart
- h. Two Lab Coats with Acu-MDx Logo on the front Pocket

5. Consumables

- a. Material required for specimen collection e.g. evacuated tubes, syringes, tubes swabs etc.
- b. Vacutainers rack / specimen holding Tray
- c. Waste disposal agreement with approved govt. agencies
- d. Urine measuring cylinder

6. Other

- a. Display hrs. of operations (Timings) outside and inside of the premises with telephone number (Preferably mobile number).
- b. Water dispenser with disposable glasses
- c. Suggestion box